

ROUTINE CHEMISTRIES	ROUTINE CHEMISTRIES	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
	ALBUMIN	70125	4 x 40 mL	800		30 days	14 days	BCG	15-25 °C
	BILIRUBIN DIRECT	70131	3 x 40 mL+1x20mL	700		15 days	9 days	DIAZO-COUPLING	15-25 °C
	BILIRUBIN TOTAL	70132	3 x 40 mL+1x20mL	700		30 days	14 days	DIAZO-COUPLING	15-25 °C
	CALCIUM	70133	4 x 40 mL	800		30 days	14 days	ARSENAZO III DYE	15-25 °C
	CREATININE	70136	4 x 40 mL	800		15 days	5 days	KINETIC ALKALINE PICRATE	2-8 °C
	GLUCOSE	70138	4 x 40 mL	800		30 days	14 days	HEXOKINASE, UV	2-8 °C
	PHOSPHOROUS	70141	4 x 40 mL	800		30 days	30 days	PHOSPHOMOLIBDATE, UV	15-25 °C
	MAGNESIUM	70143	4 x 40 mL	800		15 days	10 days	ARSENAZO I DYE	15-25 °C
	TOTAL PROTEIN	70144	4 x 40 mL	800		20 days	7 days	BUIRET	15-25 °C
	UREA NITROGEN (BUN)	70141	4 x 40 mL	800		30 days	Daily	UREASE W GLDH	15-25 °C
	URIC ACID	70147	4 x 40 mL	800		30 days	14 days	UREASE w GLDH	2-8 °C

FOLLOW THE LINKS FOR MORE ASSAYS



ENZYME ASSAYS

LIPID ASSAYS

ELECTROLYTE ASSAYS

ANEMIA ASSAYS

DIABETIC ASSAYS

CONTROLS AND CALIBRATORS

* TOTAL PRECISION is calculated @ the highest level tested

**

AMS LIASYS CLINICAL CHEMISTRY SERIES

1(866) 419-7839 (Toll Free)

ENZYME ASSAYS

ENZYME ASSAYS	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
ALP	70121	3 x 40mL+ 1x20mL	700		10 days	Not Required	IFCC	2-8 °C
ALT	70127	4 x 40 mL	800		30 days	Not Required	MODIFIED IFCC	2-8 °C
AMYLASE	70128	4 x 40 mL	800		30 days	14 days	CNPG3	2-8 °C
AST	70129	4 x 40 mL	800		30 days	14 days	MODIFIED IFCC	2-8 °C
CK	70135	3 x 40 mL+1x20mL	700		30 days	Not Required	MODIFIED SZASZ	2-8 °C
GGT	70137	3 x 40 mL+1x20mL	700		21 days	Not Required	SZASZ	2-8 °C
LDH	70142	3 x 40 mL+1x20mL	700		16 days	Not Required	MODIFIED WACKER (L-P)	2-8 °C
LIPASE**								

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

LIPID ASSAYS

ELECTROLYTE ASSAYS

ANEMIA ASSAYS

DIABETIC ASSAYS

CONTROLS AND CALIBRATORS

* TOTAL PRECISION is calculated @ the highest level tested
 ** IN VALIDATION

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
 1(866) 419-7839 (Toll Free)

LIPID ASSAYS

LIPID ASSAYS	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
HDL DIRECT	70139	3 x 40mL+1 x20mL	700		30 days	10 days	IMMUNOINHIBITION	2-8°C
LDL DIRECT	70142	3 x 40mL+1 x20mL	700		16 days	10 days		2-8°C
CHOLESTEROL	70134	4 x 40 mL	800		30 days	14 days	ENZYMATIC	2-8°C
TRIGLYCERIDES	70145	4 x 40 mL	800		30 days	14 days	ENZYMATIC - GPO	2-8°C
LIPOPROTEIN (a)**								
Apo A1 **								
Apo B **								

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

ENZYME ASSAYS

ELECTROLYTE ASSAYS

ANEMIA ASSAYS

DIABETIC ASSAYS

CONTROLS AND CALIBRATORS

* TOTAL PRECISION is calculated @ the highest level tested
 ** IN VALIDATION

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
 1(866) 419-7839 (Toll Free)

ELECTROLYTE ASSAYS

ELECTROLYTE ASSAYS	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
CO2	70150	4 x 40 mL	800		20 days	20 days	ENZYMATIC	2-8 °C
CHLORIDE	70131	1 Electrode			6 Months	Daily	ISE	25 °C
POTASSIUM	70132	1 Electrode			6 Months	Daily	ISE	25 °C
SODIUM	70133	1 Electrode			6 Months	Daily	ISE	25 °C
LITHIUM**								

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

ENZYME ASSAYS

LIPID ASSAYS

ANEMIA ASSAYS

DIABETIC ASSAYS

CONTROLS AND CALIBRATORS

* TOTAL PRECISION is calculated @ the highest level tested
 ** IN DEVELOPMENT

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
 1(866) 419-7839 (Toll Free)

Experience The Difference:

AMS LIASYS 450 TEST MENU

ANEMIA ASSAYS	ANEMIA ASSAYS	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
	IRON	70140	4 x 40 mL	800					2-8 °C
	FERRITIN								
	TIBC DIRECT								
	TRANFERRIN								

* TOTAL PRECISION is calculated @ the highest level tested
 ** IN VALIDATION

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

ENZYME ASSAYS

LIPID ASSAYS

ELECTROLYTE ASSAYS

DIABETIC ASSAYS

CONTROLS AND CALIBRATORS

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
 1(81010) 419-7839 (Toll Free)

Experience The Difference:

AMS LIASYS 450 TEST MENU

DIABETIC ASSAYS	DIABETIC ASSAYS	PART NUMBER	PRESENTATION	TESTS PER KIT	TOTAL PRECISION*	ON-BOARD STABILITY	CALIBRATION FREQUENCY	METHODS	STORAGE TEMPERATURE
	HEMOGLOBIN A1C**								15-25 °C
	MICROALBUMIN**								15-25 °C

* TOTAL PRECISION is calculated @ the highest level tested
 ** IN VALIDATION

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

ENZYME ASSAYS

LIPID ASSAYS

ELECTROLYTE ASSAYS

ANEMIA ASSAYS

CONTROLS AND CALIBRATORS

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
 1(866) 419-7839 (Toll Free)

CONTROL AND CALIBRATORS	CONTROLS AND CALIBRATORS	PART NUMBER	PRESENTATION	ON-BOARD STABILITY	CALIBRATION FREQUENCY	STORAGE TEMPERATURE
	CONTROL NORMAL	60119				15-25 °C
	CONTROL ABNORMAL	60120				15-25 °C
	CHEMISTRY CALIBRATOR	60121				15-25 °C
	LIPID CONTROL NORMAL	60122				15-25 °C
	LIPID CONTROL ABNORMAL	60123				2-8 °C
	LIPID CALIBRATOR	60148				2-8 °C
	HEMOGLOBIN A1C					

FOLLOW THE LINKS FOR MORE ASSAYS



ROUTINE CHEMISTRIES

ENZYME ASSAYS

LIPID ASSAYS

ELECTROLYTE ASSAYS

ANEMIA ASSAYS

DIABETIC ASSAYS

RETURN TO PREVIOUS PAGE



AMS LIASYS CLINICAL CHEMISTRY SERIES
1(866) 419-7839 (Toll Free)